



Certificate of Traceability

Level I Laboratory Certification

This certifies that LabCorp Central Laboratory, using Roche Cobas c503 has participated in and successfully completed NGSP Level I Laboratory certification and is traceable to the **Diabetes Control and Complications Trial Reference method**. The comparison was performed with: University of Missouri SRL#12

The system evaluated was:

Instrument: Cobas c503	Calibrator Lot: 671824	Application: (HbA1c/Hb)*91.5+2.15
Reagent Lot: 744199, 747684, 828136	Calibrator Assigned Values: 0.155, 0.360, 0.686, 0.940, 1.29, 1.83 mmol/L HbA1c; 1.48, 7.97 mmol/L Hb	

Date of Certification: January 1, 2025

Certification Expires: January 1, 2026

David Searls

NGSP Steering Committee Chair

Randa R Little PhD.

NGSP Network Coordinator

Sharon Amulfo

SRL director/ supervisor